4<sup>th</sup> HEPATITIS C
TECHNICAL ADVISORY
GROUP
TAG Meeting

# EVALUATION OF THE DIAGNOSTIC PERFORMANCE OF THE XPERT® FINGERSTICK HCV VIRAL LOAD ASSAY

Maia Alkhazashvili NCDC, Georgia

# Evaluation of the diagnostic performance of the Xpert® Fingerstick HCV Viral Load (VL) Assay

#### The Goal

To evaluate the performance of the Xpert® Fingerstick HCV VL assay in the diagnosis of HCV in different populations and laboratory settings in high-income countries and LMICs using the Abbott RealTime HCV assay as a reference test

### The Objectives of the HCV FS study


### Primary objective

 To evaluate the sensitivity, specificity and quantitation of the Xpert® Fingerstick HCV VL assay for the detection of HCV in capillary and venous whole blood

### Secondary objective

 To compare the sensitivity, specificity and quantitation of the Xpert® Fingerstick HCV VL assay in capillary and venous whole blood to that of CE IVD Xpert® HCV VL test in plasma.


### **Xpert Fingerstick HCV VL assay: workflow**



Result: ≤ 60 min

### Study enrolment overview

Between August 2017 and March 2018 we enrolled 310 participants at Harm Reduction site "Hepa plus"



## Participant demographic description

# of participants	300			
Age, years	44.8 (22-79)			
Female sex, %	15% (46/300)			
Other infections				
HIV infection	0.3% (1/300)			
Currently on antiretroviral therapy	0% (0/300)			
HBV infection	2% (7/300)			
Population groups				
HCV seropositives	66% (199/300)			
HCV risk	33% (100/300)			
Known HCV infection	0.3% (1/300)			
Risk factor exposure				
Exposure to unsafe medical procedures	37% (110/300)			
Injection of non-prescription drugs	97% (292/300)			
Positive HIV test result	0% (0/300)			
Born to HCV-positive mother	0% (0/300)			
Anti-HCV treatment within past 12 months				
IFN-only	1% (2/300)			
IFN/DAA	14% (41/300)			
DAA-only	0% (0/300)			

# Genotype distribution in HCV RNA positive participants

Genotype	%
GT1	42% (62/149)
GT2	4% (6/149)
GT3	30% (45/149)
GT4	0.7% (1/149)
GT5	0% (0/149)
GT6	0% (0/149)
Mixed	19% (28/149)
NA	5% (7/149)

# Accuracy of the Xpert Fingerstick HCV assay for detection of hepatitis C infection

	Xpert Fingerstick HCV test vs Abbott RealTime HCV test		Xpert Fingerstick HCV test vs CE-IVD Xpert® HCV Viral Load	
	Capillary blood	Venous blood	Capillary blood	Venous blood
Nr of participants	287	300	287	300
Sensitivity	95.8% (137/143)	96% (143/149)	93.2% (137/147)	93.5% (143/153)
Specificity	100% (144/144)	100% (151/151)	100% (140/140)	100% (147/147)
PPV	100%	100%	100%	100%
NPV	96%	96.2%	93.3%	93.6%
Accuracy	97.9%	98%	96.5%	96.7%

# Rate of indeterminate results of Xpert® HCV VL FS assay


 Out of 300 capillary blood samples tested by the Xpert® HCV VL FS assay, 13 (4.3%) gave indeterminate (no result, error or invalid) results


	Capillary blood	Venous blood
Number of tests performed	300	311
No Result	2 (0.7%)	2 (0.6%)
Error	8 (2.7%)	8 (2.6%)
Invalid	3 (1%)	1 (0.3%)
Total	13 (4.3%)	11 (3.5%)

# Performance of Xpert FS HCV VL assay from capillary blood

#### **Bland-Altman analysis**

CI 95 %





#### CONCLUSION

- This study was conducted to evaluate the performance characteristics of the Xpert® HCV VL FS assay for detection and quantification of HCV RNA measured against two CE-marked assays widely used for HCV diagnosis and monitoring of treatment success.
- The new Xpert® HCV VL FS assay can be performed directly from a drop of capillary blood without any additional sample preparation step
- Results of the study show that while the Xpert® HCV VL FS assay is more easy to use and has a higher potential to be utilized in decentralized settings, its performance is comparable to that of currently used standard of care HCV viral load tests.
- Results obtained in Georgia demonstrated the 95.8% sensitivity and 100% specificity of the Xpert® HCV VL FS assay in capillary blood.

### Acknowledgement

- Study participants
- Study sites
  - NCDC, Lugar Center
  - HRS, Hepa plus
- FIND
- Funders
  - Government of Netherlands
  - UNITAID
- Cepheid