5th HEPATITIS C
TECHNICAL ADVISORY
GROUP
TAG Meeting

BLOOD BANK SCREENING QUALITY ASSURANCE: CURRENT PRACTICES AND FUTURE DIRECTIONS

Nazibrola Chitadze

National Center for Disease Control and Public Health Lugar Center for Public Health Research

Outlines

- Quality assurance
- Criteria for inclusion
- EQA testing at LC
- Results
- Challenges
- Future directions


Quality assurance

 Quality systems are crucial for the overall effectiveness of all aspects of the screening programme and in assuring the quality, safety and efficacy of all blood and blood products



External quality assessment

 Assessment is a critical aspect of laboratory quality management, and it can be conducted in several ways



State Blood Safety Program



 Since 2019 - 16 out of 21 Blood Banks participate in the State Blood Safety Program

Criteria for inclusion in the State Blood Safety Program

- Possession of a license
- Required testing methods for HIV, HBV, HCV by EIA;
 - ✓ For Syphilis by TPHA, EIA or RPR
- ✓ ≥1700 donations per year;
- Annual 10% increase of unpaid donation
- ✓ At least 1000 donations unpaid donation per year (30% repeat donations)
- Provide HCV screening positive samples to Lugar center for HCVcoreAg confirmation

State Blood Safety Program requirements

- 2011 All Blood Banks, involved in the State Blood Safety Program are required to participate in the EQA
 - Aliquots from 5% of all donations are randomly selected from each Blood Bank enrolled in the State Program and rechecked by Lugar Center as part of EQA
- Since 2015, all Blood Banks, participating in the State Blood Safety Program, are required to perform international provider's Proficiency Testing (PT) quarterly

EQA confirmatory testing at Lugar Center

- Lugar Center provides regular EQA confirmatory testing for blood banks twice in a year;
- At Lugar Center samples are registered, re-labled and rechecked using Chemiluminescent microparticle immunoassay (CMIA) for HIV, HCV, HBV and Syphilis
- Discrepancies are confirmed by:
 - HIV, HCV, Syphilis Line Immunoassay
 - HBs Ag HBs Ag confirmatory (CMIA)





Retesting results of 2018 samples

In total 2937 aliquots were received from different Blood Banks participating in the state program

Confirmatory testing period	# of participating banks	# of samples	all types of discrepancy
2018 1st 6 month	14 (15-1)	1444 (1480-36)	18 (1.2%)
2018 2 nd 6 month	15	1493	19 (1.3%)

Discrepancies of 2018 samples

Discrepancies	1st 6 month # / %	2nd 6 month # / %
Testing results discrepancies between LC and the BBs	18 /1.2	18 / 1.3
Inappropriate storage and shipping (spilled samples)	36 / 2.4	0
Mislabeling	0	1 / 0.06
Concordance	1326 / 99	1460 / 98.7

Newly collected samples are requested from donors with discrepancies samples

Challenges

 Lack of standardized testing, technologies and SOPs

- No EQA are established for the Blood Banks not participating in the State Blood Safety Program.
 - Imparts quality patient care

Future directions

- Help in developing standards and guidelines for technologies and testing strategies
- Preparation, packaging and distribution of panel samples
 - Receiving, analyzing results and preparing evaluation reports
 - Provide necessary feedback, recommendations
- Involve all banks in EQA, who are not participating in state programs

Acknowledgements

 NCDC Lugar Center laboratory team - for testing and data analysis

 Blood Banks participating in the State Blood Safety Program

